



1. Color differentiation between a coated and an uncoated tube

The interior zone of a tube should be checked whether it is coated or uncoated. As shown on the photos the coated surface is slightly yellow in color whereas the uncoated surface is silver in color. For the investigations a **SPECTRO-3-30-DIL** was used and a proper differentiation between the coated and uncoated surface is possible. The color sensor was arranged at a distance of approximately 30 mm to the object.

